


Certificate of Analysis


Product	EXIGO EOS REAGENT, RFID (1.9 L)
Product No	1504502
Lot No	2007-381
Manufacturing Date	2020-07-28
Expiration Date	2022-07-28

Test	Specification
pH	8.70 – 9.30
Osmolality (mOsm/kg)	125 - 145
Conductivity (mS/cm)	6.80 – 7.60
EOS Performance *	Passes test
Blank EOS	≤ 0.1

* Performance test using calibrated Blood cell counter at Boule Medical reference laboratory.
Actual values meet specifications.

Date: 2020-08-05


Caroline Forslin
Quality Assurance



This product is intended for Laboratory use only unless otherwise specifically stated.
All information in this document is property of Boule Medical AB.

